

Desinfeksjonsmidler til <i>ultralydapparatet</i> som produsentene har bekreftet kan benyttes uten å skade deres utstyr - september 2022			Desinfeksjonsmidler til teknisk bruk <i>Bruksområder, effekt og virketider for desinfeksjonsmidler</i> Infeksjonskontroll.no og <i>desinfeksjonskluter</i> Infeksjonskontroll.no		
Produsent av ultralydprober	Varemerker	Leverandør	Pereddiksyre	Klordioksid	Bruk av alkoholer
Siemens Healthineers	ACUSON Sequoia Ultrasound System	Siemens Healthineers			
BK Medical	bkSpecto	KeboMed AS		Tristel Trio Wipe, Tristel Duo ULT	Desinfeksjonssprit 70 - 75 % på skjerm
FUJIFILM SonoSite	SonoSite	Secma AS		Tristel Trio Wipe, Tristel Duo ULT	
GE Healthcare, GE Vingmed Ultrasound	GE Healthcare	GE Healthcare Norge AS, GE Vingmed Ultrasound		Tristel Trio Wipe, Tristel Duo ULT	
Philips	Philips ATL	Philips Norge AS			70 % isopropylalkohol (IPA) (ikke godkjent for berøringsskjermer)
Hitachi – Aloka	Aloka	Scan-Med AS Norway	Rely+On PeraSafe		
Canon Medical	Canon Aplio, Xario, Aviamo	Tromp Medical AS	Rely+On PeraSafe	Tristel Trio Wipe, Tristel Duo ULT	
Samsung Healthcare	Samsung	Inter Medical AS		Tristel Trio Wipe, Tristel Duo ULT,	
Natus Neuro	Natus® InVisus™ Ultrasound Sonara	Timik AS		Tristel Trio Wipe, Tristel Duo ULT	

