Oslo universitetssykehus

Desinfeksjonsmidler som produsentene har bekreftet kan benyttes uten å skade deres *ultralydutstyr* - september 2022			**Desinfeksjonsmidler til teknisk bruk** *Bruksområder, effekt og virketider for desinfeksjonsmidler Infeksjonskontroll.no og desinfeksjonskluter Infeksjonskontroll.no*		
Produsent av ultralydprober	**Varemerker**	**Leverandør**	**Pereddiksyre**	**Klordioksid**	**Bruk av alkoholer**
Siemens Healthineers	**ACUSON Sequoia Ultrasound System**	**Siemens Healthineers**			
BK Medical	**bkSpecto**	**KeboMed AS**		**Tristel Trio Wipe, Tristel Duo ULT**	**Desinfeksjonssprit 70 - 75 % på skjerm**
FUJIFILM SonoSite	**SonoSite**	**Secma AS**		**Tristel Trio Wipe, Tristel Duo ULT**	
GE Healthcare, GE Vingmed Ultrasound	**GE Healthcare**	**GE Healthcare Norge AS, GE Vingmed Ultrasound**		**Tristel Trio Wipe, Tristel Duo ULT**	
Philips	**Philips ATL**	**Philips Norge AS**			**70 % isopropylalkohol (IPA) (ikke godkjent for berøringsskjermer)**
Hitachi – Aloka	**Aloka**	**Scan-Med AS Norway**	**Rely+On PeraSafe**		
Canon Medical	**Canon Aplio, Xario, Aviamo**	**Tromp Medical AS**	**Rely+On PeraSafe**	**Tristel Trio Wipe, Tristel Duo ULT**	
Samsung Healthcare	**Samsung**	**Inter Medical AS**		**Tristel Trio Wipe, Tristel Duo ULT,**	
Natus Neuro	**Natus® InVisus™ Ultrasound Sonara**	**Timik AS**		**Tristel Trio Wipe, Tristel Duo ULT**	